SURGICAL KIT



Epikut Long Surgical Kit – KCSEL 01

The Epikut Long Surgical Kit is intended for specialized procedures, which must be performed by qualified professionals. The use of the product and surgical techniques are inherent to the professional's training. The use of the product must be performed in a surgical environment and under adequate conditions for the patient's health and safety.



PROUDCT DESCRITION

The Epikut Long Surgical Kit is composed of materials required to install long implants from the Epikut and Epikut Plus lines.

Lance Cutter: Used for marking the site where the implant will be installed, breaking the cortical bone and facilitating the insertion of other cutters.

Conical Cutters: Used in sequence for the progressive diameter enlargement of the drilling until the final dimension is adequate for the implant installation following the surgical planning.

Direction Indicator: Its purpose is to help in the correct drilling of the surgical alveolus.

Key/Fixator: Its purpose is to enable the installation of dental implants and prosthetic components, through the aid in the surgical procedure.

Depth Probe: Its purpose is to help in the correct drilling of the surgical alveolus.

Torque Meter: Its purpose is to enable the installation of dental implants and prosthetic components, through the aid in the surgical procedure.

Box: Support for storing, transporting and sterilizing all items of the kit.

INDICATIONS OF USE

The Epikut Long Surgical Kit is indicated for the installation of implants in the maxilla and mandible in surgical procedures with delayed and immediate loading and for single implants and multiple rehabilitations.

OPERATION PRINCIPLE

The instruments contained in the Epikut Long Surgical Kit are based on mechanical action. All instruments are indicated for use in the placement of long implants of the line Epikut and Epikut Plus and must be used following the appropriate dental techniques.

HOW TO USE

The Epikut Long Surgical Kit should be used in accordance with the surgical planning programmed by a qualified dentistry professional. Each instrument included in the kit has a way of use, which can be accessed on the website: www.sinimplantsystem.com

The Epikut Long Surgical Kit are intended for specialized procedures, which must be performed by qualified professionals in implantology. The use of the product must be performed in a surgical environment and in proper conditions for the health and safety of the patient.

PRECAUTIONS

The product should only be used by qualified dental professionals who already have all the scientific information necessary for the correct use of the product;

Always perform cleaning and sterilization as recommended before the surgical procedure;

Before drilling, make sure the cutter fits into the contraangle and the motor is adjusted for rotation, torque and irrigation;

During drilling the pressure must not be excessive, and intermittent movements must be made with constant irrigation;



The drills are not capable of sharpening and using them without cutting can generate undue bone heating, compromising the success of the procedure;

The use of the drills or the inappropriate cutter sequence can compromise the performance of the implant resulting in system failures, such as implant loss or fracture;

Care should be taken in cases of patients who show signs of allergy or hypersensitivity to stainless steel;

For drills, a maximum use of 20 to 30 perforations is recommended, as follows: 20 perforations in high density bones and 30 perforations in low density bones;

Do not stick labels, adhesive tapes, write or mark the surface of the product.

RECOMMENDATIONS

Before each procedure, check the condition of the instruments, always respecting their useful life. It is necessary to replace the instruments in case of damage, erased markings, compromised sharpening, deformations and wear.

CONTRAINDICATIONS

Use of drills without irrigation, which can cause bone necrosis. Use for purposes other than installing a Epikut implant.

SIDE EFFECTS

It will not occur as long as the surgical planning and handling is done according to the instructions for use.

WARNING

Do not use the instruments if you notice cracks, wear or oxidation/corrosion points. This may cause problems in the functioning of the products, installation of the implants and in the post-operative period. Some items present natural wear and tear generated by use, such as cutters, and must be replaced whenever the professional identifies loss of cutting capacity or precision of these products, as they can interfere with the final result of the treatment.

TRACEABILITY

All S.I.N. - Implant System products have sequential batches that allow traceability, thus promoting greater safety to the professional skilled in the procedure. Through this lot number it is possible to know all product history from the manufacturing process to the time of distribution.

STORAGE

The Epikut Long Surgical Kit should be stored in a cool, dry place at a maximum temperature of 35°C and protected from direct sunlight.

HANDLING

Once sterilized, the instruments should be handled only in a sterile environment by properly attired professionals and in appropriate clothing at the time of surgery to install implants.

DISPOSAL OF MATERIAL

The disposal of materials should comply with local hospital regulations and applicable local laws.

TRANSPORTATION

The Epikut Long Surgical Kit must be transported adequately to avoid falling and stored under a maximum temperature of 35°C, protected from heat and moisture. Carriage must be carried out in its original packaging.

COMPLEMENTARY INFORMATIONS

Multiple use product. Reprocessing allowed. Refer to the cleaning and sterilization conditions contained in these instructions. In case of an incident caused by the product, the professional must immediately inform the manufacturer. If you need the printed version of this instruction for use, without any cost, please request by e-mail to sin@sinimplante.com.br or call to 0800 770 8290 will receive until 7 days calendar.



CLEANING INSTRUCTIONS

- 1. Remove manually all surgical instruments from the kit. Remove the kit box parts (lid, tray and bottom);
- 2. Prepare the enzymatic detergent, according to manufacturer's recommendation;
- Immerse the trays into the prepared detergent solution and keep in contact for at least 5 minutes, then using a soft bristle brush, scrub the parts to remove organic matter from the products;
- Remove trays from detergent solution and rinse with tap water for 1 minute, repeat the rinse for two more times, a total of three rinses of 1 minute each;
- 5. Visual inspection of each part for cleaning process residue or organic waste from product use;
- If residue is detected in the product, repeat the cleaning process until the residue is completely removed;
- 7. Dry with a soft, clean, dry cloth or disposable paper.

RECOMMENDATIONS

- a. Use the proper PPEs (gloves, masks, goggles, caps, etc.);
- b. Start the cleaning right after the surgical use;
- c. Never let the instruments dry with organic waste after the surgical use;
- d. Never let the instrument dry naturally after cleaning;
- Never use saline solutions, include sodium hypochlorite, disinfectant, hydrogen peroxide or alcohol to cleaning or rinsing the surgical instruments and kits;
- f. Never use steel wool and abrasive products, so that the instruments are not damaged;
- g. Do not stack the instruments in lots to avoid the deformation of smaller and delicate pieces.

TORQUE RATCHET CLEANING INSTRUCTIONS

Cleaning must be carried out immediately after using the torque wrench. For cleaning, the torque wrench must be disassembled, for this it is not necessary to use tools.

- 1. Pull the steering inverter rod back;
- 2. Remove the ratchet from the socket with the head;
- 3. Rotate the fixing door counterclockwise;
- 4. Remove the central axis from the torque wrench;
- 5. Remove the torque graduated rod;

- 6. Prepare the enzymatic detergent according to the manufacturer's instructions.
- 7. Dip all parts of the product in the prepared detergent solution and leave for at least 5 minutes, afterwards using a soft bristle brush, rub the parts to remove organic matter from the products.
- 8. Remove the parts from the detergent solution and rinse under running water for 1 minute, repeat the rinse for two more times, totaling 3 rinses of 1 minute each.
- 9. Visually inspect each part to check for residue from the cleaning process or organic residues from using the product.
- 10. If the presence of residues in the product is confirmed, repeat the cleaning process, until the total removal of residues.
- 11. Dry with a soft, clean, dry cloth or disposable paper.

STERILIZATION

Reusable product and provided non-sterile. It must be clean and sterilized in autoclave before use.

- 1. Dry all instruments before the steam sterilization cycle;
- 2. The product must be enclosed in a steam sterilizable wrap;
- Steam sterilize in cycles of 121°C at 1 ATM pressure for 30 minutes or of 134°C at 2 ATM pressure for 20 minutes. Drying time 30 minutes;
- 4. Always accommodate the case in autoclave over a plane surface and away of device walls;
- 5. Never stack objects or other cases.

RECOMMENDATIONS

- a. Sterilize the products in the same day or one day earlier the procedure;
- b. The chemical sterilization is not recommended, once some products may cause the discoloration and damages to the case;
- c. Do not use temperature higher than 60°C to drying process;
- d. Do not use dry heat stoves for sterilization of the instruments and kits from S.I.N. Implant System.

LIFE TIME

The Epikut Long Surgical Kit can be used as below depending on the proper handling, cleaning and sterilization.



Drill and Extracting Drill: 20 perforations in high density bones; 30 low density bone perforations.

Drive/Fixer, Guide Drill, Depth Rod and Torque Ratchet: Used up to 250 times.

Safe Drill Stopper: Used up to 50 times.

Box: The box can be cleaned and sterilized up to 250 times.



NON	NÃO ESTÉRIL	NON-ESTERILE	NO ESTÉRIL
Ĩ	CONSULTAR AS INSTRUÇÕES DE USO	CONSULT INSTRUCTIONS FOR USE	CONSULTE LAS INSTRUCCIONES DE USO
Ť	MANTENHA SECO	KEEP DRY	MANTÉNGALO SECO
溇	MANTENHA AO ABRIGO DO SOL	KEEP AWAY FROM SUNLIGHT	MANTÉNGALO LEJOS DE LA LUZ SOLAR
8	NÃO UTILIZAR SE A EMBALAGEM ESTIVER VIOLADA	DO NOT USE IF PACKAGE IS DAMAGED	NO LO UTILICE SI EL ENVOLTORIO ESTÁ DAÑADO
Λ	ATENÇÃO	CAUTION	PRECAUCIÓN
35°C	LIMITE SUPERIOR DE TEMPERATURA	UPPER LIMIT OF TEMPERATURE	LÍMITE SUPERIOR DE TEMPERATURA
Rx only	ATENÇÃO: A LEI FEDERAL (EUA) LIMITA A VENDA DESTE DISPOSITIVO POR OU POR ORDEM DE UM PROFISSIONAL DE SAÚDE LICENCIADO.	(USA) RESTRICTS THIS DEVICE TO SALE BY OR ON	PRECAUCIÓN: LAS LEYES FEDERALES (USA) RESTRINGEN LA VENTA DE ESTE DISPOSITIVO POR O EN EL ORDEN DE UN PROFESIONAL DE LA SALUD LICENCIADO.
	FABRICANTE	MANUFACTURE	FABRICANTE
	DATA DE FABRICAÇÃO	DATE OF MANUFACTURE	FECHA DE FABRICACIÓN
REF	CÓDIGO DE REFERÊNCIA	REFERENCE CODE	CÓDIGO DE REFERENCIA
MD	DISPOSITIVO MÉDICO	MEDICAL DEVICE	DISPOSITIVO MEDICO
UDI	IDENTIFICADOR ÚNICO DO DISPOSITIVO	UNIQUE DEVICE IDENTIFIER	IDENTIFICADOR DE DISPOSITIVO ÚNICO
		UNIQUE DEVICE IDENTIFIER	
	DO DISPOSITIVO		DISPOSITIVO ÚNICO
	DO DISPOSITIVO IMPORTADOR	IMPORTER	DISPOSITIVO ÚNICO IMPORTADOR
	DO DISPOSITIVO IMPORTADOR DISTRIBUIDOR	IMPORTER	DISPOSITIVO ÚNICO IMPORTADOR DISTRIBUIDOR



DEVELOPED AND MANUFACTURED BY: S.I.N. Sistema de Implante Nacional S/A

CNPJ [Corporate Taxpayer's Registry]: 04.298.106/0001-74 Rua Soldado Ocimar Guimarães da Silva, 2445 - Vila Rio Branco CEP: 03348-060 - São Paulo - SP - Brazil

SERVICE TO PROFESSIONALS

0800 770 8290 +55 (11) 2169-3000 www.sinimplantsystem.com email: sin@sinimplante.com.br

RESPONSIBLE TECHNICIAN:

Alessio Di Risio CREA-SP (register): 5061207169

PRODUCT: Epikut Long Surgical Kit - KCSEL 01

ANVISA REGISTRATION: 80108910103